

1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product LOT	101555
Date Issued	12MAR24
Expiration Date	MAR2025

Population	% of Population*	Results (Pass/Fail)
Lymphocyte Mimics	28-36%	Pass
Monocyte Mimics	28-36%	Pass
Granulocyte Mimics	28-36%	Pass

*"% of Population" is measured by FSC and SSC

Approved by: Simon Au Quality Engineer 03 / 12 / 2024
Quality Assurance