

1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product LOT	101933
Date Issued	26JUL2024
Expiration Date	JUL2025

Population	% of Population*	Results (Pass/Fail)
Lymphocyte Mimics	28-36%	Pass
Monocyte Mimics	28-36%	Pass
Granulocyte Mimics	28-36%	Pass

*"% of Population" is measured by FSC and SSC

Approved by: Jonathan Yang Director, QA 07 / 26 / 2024
Quality Assurance