

# 1. Certificate of Analysis

<b>Scope</b>	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
<b>Product Lot</b>	101733
<b>Date Issue</b>	MAY2024
<b>Expiration Date</b>	MAY2025

## Section A: Reactivity

Species	Reactivity Confirmed (Yes/No)
Mouse	Yes
Rat	Yes
Hamster	Yes
Rabbit	Yes
Human	Yes

## Section B: MFI\*

Population	MFI Specification	Results (Pass/Fail)
Mouse Positive (stained using Mouse IgG1 Anti-human CD3 FITC; clone-OKT3)	$\geq 400,000$	Pass
Rat Positive (stained using Rat IgG1 Anti-mouse CD25 FITC clone PC61.5)	$\geq 400,000$	Pass
Hamster Positive (stained using Armenian Hamster Anti-mouse CD3e FITC; clone 145-2C11)	$\geq 400,000$	Pass
Rabbit Positive (stained using Rabbit anti-human kappa FITC)	$\geq 400,000$	Pass
Human Positive (stained with Human IgG1 FITC (clone QA16A12))	$\geq 400,000$	Pass

\*when analyzed using Cytex Aurora 5 laser Cytometer under recommended gain settings

Approved by: Jonathan Yang Director, Quality Assurance 05 / 06 / 2024

Slingshot Biosciences Internal Part Number: 11-0623v1