

Slingshot Biosciences
SpectraComp eGFP (SSB-23-A) Certificate of Analysis.

COA-16
Version: 2.0

1. Certificate of Analysis

| | |
|------------------------|--|
| Scope | The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures. |
| Product Lot | 102138 |
| Date Issued | 18OCT2024 |
| Expiration Date | OCT2025 |

| Population | MFI specification | Results (Pass/Fail) |
|---------------------------|-------------------|---------------------|
| SpectraComp eGFP Negative | <3000 | Pass |
| SpectraComp eGFP Low | 10k - 30k | Pass |
| SpectraComp eGFP Medium | 100k - 250k | Pass |
| SpectraComp eGFP High | 1M - 2.5M | Pass |

***when analyzed using Cytek Aurora 5 laser Cytometer under recommended gain settings**

Approved by:  Signed by: **Jonathan Yang** Director, QA/Regulatory 10/18/2024
B81B0F0EDC0144B...

Quality Assurance

Slingshot Biosciences Internal Part Number: 11-0630v1