Version: 5.0

COA-3

1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product LOT	101583
Date Issued	25Mar2024
Expiration Date*	Mar2025

^{*}Product has a shelf life of six months from the date of thaw. Follow the Date of Expiration if it occurs before shelf life expiration.

Section A: Staining Dyes

Stain	Staining Confirmed (Yes/No)
7AAD	Yes
Ghost	Yes

Section B: MFI*

Population	MFI specification	Results (Pass/Fail)
Negative for 7AAD	<u>≤</u> 500	Pass
Positive for 7AAD	<u>≥</u> 20,000	Pass
Negative for Ghost	<u>≤</u> 15,000	Pass
Positive for Ghost	<u>≥</u> 200,000	Pass

^{*}when analyzed using on Cytek Aurora 5 laser Cytometer under recommended gain settings

Jonathan Yang Director of Quality Assurance 03 / 25 / 2024 Approved by:

Quality Assurance

Slingshot Biosciences Internal Part Number: SSB-11-0552