

# 1. Certificate of Analysis

<b>Scope</b>	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
<b>Product LOT</b>	102121
<b>Date Issued</b>	27SEP2024
<b>Expiration Date*</b>	SEP2025

\*Product has a shelf life of six months from the date of thaw. Follow the Date of Expiration if it occurs before shelf life expiration.


## Section A: Staining Dyes

Stain	Staining Confirmed (Yes/No)
7AAD	Yes
Ghost	Yes

## Section B: MFI\*

Population	MFI specification	Results (Pass/Fail)
Negative for 7AAD	$\leq 500$	Pass
Positive for 7AAD	$\geq 20,000$	Pass
Negative for Ghost	$\leq 15,000$	Pass
Positive for Ghost	$\geq 200,000$	Pass

\*when analyzed using on Cytex Aurora 5 laser Cytometer under recommended gain settings

Signed by:  
  
 Approved by: Jonathan Yang Director, QA/Regulatory 9/27/2024  
B81B0F0EDC0144B...

Quality Assurance

Slingshot Biosciences Internal Part Number: SSB-11-0552