

1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product Lot	102291
Date Issued	JAN 2025
Expiration Date	JAN 2026

Section A: Reactivity

Species	Reactivity Confirmed (Yes/No)
Mouse	Yes
Rat	Yes
Hamster	Yes

Section B: MFI*

Population	MFI specification	Results (Pass/Fail)
Mouse Positive (stained using Mouse IgG1 Anti-human CD8 FITC; clone-SK1)	$\geq 400,000$	Pass
Rat Positive (stained using Rat IgG1 Anti-mouse CD25 FITC clone PC61.5)	$\geq 400,000$	Pass
Hamster Positive (stained using Armenian Hamster Anti-mouse CD3e FITC; clone 145-2C11)	$\geq 400,000$	Pass

*when analyzed using Cytek Aurora 5 laser Cytometer under recommended gain settings

Jonathan Yang 01 / 28 / 2025

Approved by: _____

Quality Assurance

Slingshot Biosciences Internal Part Number: 11-0601v1