

Slingshot Biosciences
SpectraComp Certificate of Analysis (SSB-05-A/B)

COA-7
Version: 5.0

1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product Lot	102239
Date Issued	23DEC2024
Expiration Date	03DEC2025


Section A: Reactivity

Species	Reactivity Confirmed (Yes/No)
Mouse	Yes
Rat	Yes
Hamster	Yes

Section B: MFI*

Population	MFI specification	Pass/Fail
Mouse Positive (stained using Mouse IgG1 Anti-human CD8 FITC; clone-SK1)	$\geq 400,000$	Pass
Rat Positive (stained using Rat IgG1 Anti-mouse CD25 FITC clone PC61.5)	$\geq 400,000$	Pass
Hamster Positive (stained using Armenian Hamster Anti-mouse CD3e FITC; clone 145-2C11)	$\geq 400,000$	Pass

*when analyzed using Cytex Aurora 5 laser Cytometer under recommended gain settings

Approved by:  Signed by: *Simon Au* Quality Engineer 12/23/2024
F2D266B88A494B3...

Quality Assurance

Slingshot Biosciences Internal Part Number: 11-0530v2