

1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product LOT	102137
Date Issued	04OCT2024
Date of Expiration*	OCT2025

*Product has a shelf life of six months from the date of thaw. Follow the Date of Expiration if it occurs before shelf life expiration.


Section A: Staining Dyes

Stain	Staining Confirmed (Yes/No)
7AAD	Yes
Ghost	Yes

Section B: MFI**

Population	MFI specification	Results (Pass/Fail)
Positive for 7AAD	$\geq 20,000$	Pass
Positive for Ghost	$\geq 200,000$	Pass

**when analyzed using on Cytek Aurora 5 laser Cytometer under recommended gain settings

Approved by:  Signed by: Jonathan Yang Director, QA/Regulatory 10/4/2024
B81B0E0EDC0144B

Quality Assurance