

Slingshot Biosciences  
SpectraComp eGFP (SSB-23-A) Certificate of Analysis.

COA-16  
Version: 2.0


1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product Lot	102560
Date Issued	JUL2025
Expiration Date	JUL2026

Population	MFI specification	Results (Pass/Fail)
SpectraComp eGFP Negative	<3000	Pass
SpectraComp eGFP Low	10k - 30k	Pass
SpectraComp eGFP Medium	100k - 250k	Pass
SpectraComp eGFP High	1M - 2.5M	Pass

\*when analyzed using Cytex Aurora 5 laser Cytometer under recommended gain settings


Signed by:



Quality Engineer

7/29/2025

Approved by:



Quality Assurance

Slingshot Biosciences Internal Part Number: 11-0630v1