

1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product LOT	102435
Date Issued	APR2025
Expiration Date	23APR2026

Population	% of Population*	Results (Pass/Fail)
Lymphocyte Mimics	28-36%	Pass
Monocyte Mimics	28-36%	Pass
Granulocyte Mimics	28-36%	Pass

*"% of Population" is measured by FSC and SSC

Approved by:  Signed by:
Jonathan Yang
B81B0F0EDC0144B... Director, Quality and Regulatory 5/23/2025

Quality Assurance