

# 1. Certificate of Analysis

<b>Scope</b>	The product details below have not been cleared by the FDA for <i>In Vitro</i> Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
<b>Product Lot</b>	102784
<b>Date Issued</b>	MAR2026
<b>Expiration Date</b>	JUN2027


## Section A: Reactivity

Species	Reactivity Confirmed (Yes/No)	
Mouse	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No
Rat	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No
Hamster	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No

## Section B: MFI\*

Population	MFI Specification	Pass/Fail	
Mouse Positive (stained using Mouse IgG1 Anti-human CD8 FITC; clone-SK1)	≥400,000	<input checked="" type="checkbox"/> Pass	<input type="checkbox"/> Fail
Rat Positive (stained using Rat IgG1 Anti-mouse CD25 FITC clone PC61.5)	≥400,000	<input checked="" type="checkbox"/> Pass	<input type="checkbox"/> Fail
Hamster Positive (stained using Armenian Hamster Anti-mouse CD3e FITC; clone 145-2C11)	≥400,000	<input checked="" type="checkbox"/> Pass	<input type="checkbox"/> Fail

\* when analyzed using Cytex Aurora 5 Laser Cytometer under recommended gain settings

Signed by:  
 Approved by:  3/6/2026  
 D62A2FF04F7F446...

Quality Assurance